

**Certificate of Calibration
Fluke Calibration
Wuhu Laboratory**

Description:	PRESSURE MODULE	Certificate Number:	
Manufacturer:	Fluke Corporation	Date of Calibration:	22 Jan 2019
Model:	750P05	Date Due:	
Serial Number:		Temperature:	20 to 26 °C
Status:	AS-LEFT	Relative Humidity:	10 to 70 %RH
Calibration:	FULL	Pressure:	97 to 103 kPa
Procedure:	750 Module Final Master 1.1 (1 year) : 1.1	Issue Date:	22 Jan 2019

Customer:

This calibration is traceable to the International System of Units (SI) through recognized national metrology institutes (NIST, PTB, NPL, NIM, NRC, etc.), radiometric techniques, or natural physical constants. The calibration has been completed in accordance with the Fluke Quality System document QSD 111.0. This certificate applies to only the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation.

If measurement uncertainties are provided on the certificate of calibration, they are calculated in accordance with the method described in the ISO Guide to the Expression of Uncertainty in Measurement. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k such that the coverage probability corresponds to approximately 95 %.

Comments:



**Teng Yawen
Calibration Technician**

Certificate Number:

Date of Calibration: 22 Jan 2019

Standards Used

Description	Serial Number	Due-Date
Fluke Calibration 7252i 100/36 Pressure Controller/Calibrator	69295	19-Apr-2019

Quality Manuals

This calibration has been completed in accordance with:

- The Fluke Corporate Quality Manual, QSD 111.0, Revision 121, Dated July, 2017 and/or
- The Fluke 17025 Quality Manual, QSD 111.41, Revision 005, Dated Sept. 2014

Certificate Number:

Date of Calibration: 22 Jan 2019

Applied	Test Result	Error	Specification
0.000 to 30.000 psi Verification			
0.0000 psi	-0.000 psi	-0.0000 psi	0.011 psi
6.0000 psi	5.999 psi	-0.0010 psi	0.011 psi
12.0000 psi	11.999 psi	-0.0010 psi	0.011 psi
18.0000 psi	18.000 psi	0.0000 psi	0.011 psi
24.0000 psi	24.000 psi	0.0000 psi	0.011 psi
30.0000 psi	30.001 psi	0.0010 psi	0.011 psi
24.0000 psi	24.001 psi	0.0010 psi	0.011 psi
18.0000 psi	18.001 psi	0.0010 psi	0.011 psi
12.0000 psi	12.001 psi	0.0010 psi	0.011 psi
6.0000 psi	6.000 psi	0.0000 psi	0.011 psi
0.0000 psi	0.000 psi	0.0000 psi	0.011 psi
Leak Test Verification: Passed			